

|                     |                          |                      |                                 |
|---------------------|--------------------------|----------------------|---------------------------------|
| ULR Number          | ULR - CC417824000000359F | Certificate Number   | RPL-CF-2024-ET-CSRF-1105-C-2513 |
| Date of Calibration | 15.11.2024               | Recommended due date | 14.11.2025                      |

## Electro Technical Calibration Certificate

For

### Inverter ARC Welding Machine (Ammeter & Voltmeter)

Laboratory certifies that the equipment has been calibrated in accordance with a quality system registered to ISO 9001:2015 and confirms to ISO/IEC 17025:2017 and laboratory applicable procedures.

The measurement expanded uncertainty is expressed at 95% confidence level with a coverage factor equal to two ( $k=2$ ).

Laboratory measurement standards are traceable through unbroken chain of comparison to National and / or to International Standards.

Calibration of all equipment, meters used are traceable to National Standards / International Standards and their calibration are current.

This calibration report contains the results of calibration and / or other relevant data about the calibration.

This report applies only to the item identified and shall not be reproduced except in full, without specific written approval from the Laboratory.

No statement of compliance with specifications is made or implied on this report. However, the results are reviewed to establish where any measurement results exceed the manufacturers specification and communicate results by means of this report.



Lab QR Code



Certificate QR Code

| Calibration Procedure                         | As Received Conditions | As Shipped Condition |
|-----------------------------------------------|------------------------|----------------------|
| RPL-L03-DI-025-ET-004 / RPL-L03-DI-002-ET-002 | Satisfactory           | Cleaned & Calibrated |

| Calibrated By  |           | Reviewed & Approved By   Authorized Signatory |                     |
|----------------|-----------|-----------------------------------------------|---------------------|
| Identification | RPL-TM-02 | Identification                                | Jayasree Manikandan |

Page No : 01 of 05



|                            |                          |                             |                                 |
|----------------------------|--------------------------|-----------------------------|---------------------------------|
| <b>ULR Number</b>          | ULR - CC417824000000359F | <b>Certificate Number</b>   | RPL-CF-2024-ET-CSRF-1105-C-2513 |
| <b>Date of Calibration</b> | 15.11.2024               | <b>Recommended due date</b> | 14.11.2025                      |

|                                           |            |                             |            |
|-------------------------------------------|------------|-----------------------------|------------|
| <b>Date of Performance of Calibration</b> | 15.11.2024 | <b>Date of Issue</b>        | 18.11.2024 |
|                                           |            | <b>Recommended Due Date</b> | 14.11.2025 |

|                                                    |                                                                                               |                               |                       |
|----------------------------------------------------|-----------------------------------------------------------------------------------------------|-------------------------------|-----------------------|
| <b>Name &amp; Contact information of Customer:</b> | Peekay Steel Castings Private Limited<br>Plot No. 40, APIIC IP Gollapuram, Hindupur - 515211. | <b>Date of Receipt</b>        | 15.11.2024            |
|                                                    |                                                                                               | <b>CSRF Number</b>            | RPL/SRF/2024/1105     |
|                                                    |                                                                                               | <b>Location-Calibrated at</b> | Onsite                |
|                                                    |                                                                                               | <b>PO No. / PO Date</b>       | 71000480 / 13.11.2024 |

**Description of the Unit Under Calibration (UUC)**

|                    |                                                    |                          |                                 |
|--------------------|----------------------------------------------------|--------------------------|---------------------------------|
| <b>Description</b> | Inverter ARC Welding Machine (Ammeter & Voltmeter) | <b>Model</b>             | ARC601i                         |
| <b>Range</b>       | (50 to 600 A) & (22 to 44 V DC)                    | <b>Serial No</b>         | V102.00301290-231204346         |
| <b>Least Count</b> | 1V / 1A                                            | <b>Identification No</b> | PKH-WEL-26                      |
| <b>Make</b>        | ESAB-ARC601i                                       | <b>Accuracy</b>          | ±2.5% FS & ±1.5V                |
| <b>Location</b>    | Fettling                                           | <b>Operating Range</b>   | (50 to 600 A) & (22 to 44 V DC) |
| <b>Part Number</b> | PKH-TMME-1107                                      |                          |                                 |

|                                                       |                                                |
|-------------------------------------------------------|------------------------------------------------|
| <b>Condition of unit under calibration on receipt</b> | Satisfactory                                   |
| <b>Method used - Reva Phoenix Document Number</b>     | RPL-L03-DI-025-ET-004 / RPL-L03-DI-002-ET-002  |
| <b>Reference Standard (International / National)</b>  | IS 13875 – Part 1 & 2, BS EN IEC 60974-14:2018 |

|                              |                          |              |         |            |         |                   |         |
|------------------------------|--------------------------|--------------|---------|------------|---------|-------------------|---------|
| <b>Environmental Details</b> | <b>Temperature</b>       | <b>Start</b> | 25.5 °C | <b>End</b> | 25.5 °C | <b>Difference</b> | 0.2 °C  |
|                              | <b>Relative Humidity</b> |              | 58 %RH  |            | 59 %RH  |                   | 1.0 %RH |

**Details of Traceability - Master Equipment / Standard used for Calibration**

| Sl. No. | Description Master Equipment / Standard used for calibration | Metrological Traceability to National / International Standards |
|---------|--------------------------------------------------------------|-----------------------------------------------------------------|
| 1       | Multi Product Calibrator                                     | CR/PCAL/52891/1 Valid till 18.07.2025                           |
| 2       | Precision Multimeter                                         | RPL-CF-2024-ET-CSRF-0746-C-1501 Valid till 02.08.2025           |

|                       |           |                                                          |                     |
|-----------------------|-----------|----------------------------------------------------------|---------------------|
| <b>Calibrated By</b>  |           | <b>Reviewed &amp; Approved By   Authorized Signatory</b> |                     |
| <b>Identification</b> | RPL-TM-02 | <b>Identification</b>                                    | Jayasree Manikandan |

|                            |                          |                             |                                 |
|----------------------------|--------------------------|-----------------------------|---------------------------------|
| <b>ULR Number</b>          | ULR - CC417824000000359F | <b>Certificate Number</b>   | RPL-CF-2024-ET-CSRF-1105-C-2513 |
| <b>Date of Calibration</b> | 15.11.2024               | <b>Recommended due date</b> | 14.11.2025                      |

**Discipline :** Electro Technical

**Calibration Results**

|                         | A   | A   | A  | A    |
|-------------------------|-----|-----|----|------|
| AMMETER : (Input : DCV) |     |     |    |      |
| 1                       | 50  | 50  | 0  | 0.58 |
| 2                       | 100 | 100 | 0  | 0.58 |
| 3                       | 200 | 199 | -1 | 0.58 |
| 4                       | 400 | 399 | -1 | 0.59 |
| 5                       | 600 | 599 | -1 | 0.59 |
| VOLTMETER :             |     |     |    |      |
|                         | V   | V   | V  | V    |
| 1                       | 22  | 22  | 0  | 0.14 |
| 2                       | 30  | 30  | 0  | 0.14 |
| 3                       | 35  | 34  | -1 | 0.15 |
| 4                       | 40. | 39  | -1 | 0.15 |
| 5                       | 44  | 43  | -1 | 0.15 |

**Labs & Consultants Pvt. Ltd.**
**Digital Signature of Laboratory:**

REVA PHOENIX LABS AND CONSULTANTS

PRIVATE LIMITED

Date: 08/01/2025 12:05:05 +05:30

Reason: Calibration Certificate

Location: Chennai

|                       |           |                                                          |                     |
|-----------------------|-----------|----------------------------------------------------------|---------------------|
| <b>Calibrated By</b>  |           | <b>Reviewed &amp; Approved By   Authorized Signatory</b> |                     |
| <b>Identification</b> | RPL-TM-02 | <b>Identification</b>                                    | Jayasree Manikandan |

Page No : 03 of 05

|                            |                          |                             |                                 |
|----------------------------|--------------------------|-----------------------------|---------------------------------|
| <b>ULR Number</b>          | ULR - CC417824000000359F | <b>Certificate Number</b>   | RPL-CF-2024-ET-CSRF-1105-C-2513 |
| <b>Date of Calibration</b> | 15.11.2024               | <b>Recommended due date</b> | 14.11.2025                      |

### Mandatory Annexure to Calibration Certificate

- Calibration Certificate includes all the information as agreed with Customer and as necessary for the interpretation of the results and all information required by the method used, including, additions to deviations, or exclusions.
- All issued reports are retained as technical records by our Laboratory as per ISO/IEC 17025: 2017 requirements.
- Metrological traceability to National | International Standards as per ISO/IEC 17025: 2017 requirements is and Calibration reports do not contain any recommendation on Calibration interval, except where agreed.
- Laboratory is responsible for all the information provided in the report, except when information is provided by the customer. Data provided, where applicable, is clearly identified, as they can affect the validity of results.
- Measurement Uncertainty reported is at approximately 95% confidence level with  $k=2$ ; Unit of measurement result and measurement uncertainty are same as that of range selected, unless otherwise indicated in certificate.
- Calibration Certificate is issued in accordance with the terms and conditions laid down by NABL which has assessed our measurement capability and its traceability to recognized National | International standards.
- Metrological traceability of Masters to National / International Standard is established by unbroken chain of calibration.
- Calibration Certificate pertains only to the particular item(s) submitted and reported for calibration and shall not be reproduced except in full, without the written approval from our Laboratory.
- Results reported in Calibration Certificate are valid at the time of and stated conditions of measurement of UUC.
- Error(s), if any, in the Calibration Certificate, shall be brought to the notice of Laboratory within 30 days from the date of issue. Amendments to Certificate will not be entertained thereafter.
- Calibration Certificate is issued based on Customer requirements for Scientific and Industrial purposes and not meant for other purpose and shall not be produced in the court of law.
- Legal Aspects: Calibration of Weights, Density of solids, Load cell, UTM, Pressure Dead Weight Tester, Pressure Gauge. Liquid-in-glass Thermometer done by our Laboratory is meant for scientific and industrial purpose only. However, if used for commercial trading, additional recognition/ approval shall be complied as required by Department of Legal Metrology, Regulatory body, etc.
- Laboratory is not liable for any change in the reported data and performance specification on account of malfunctioning of Standards | Equipment covered by this Certificate due to damage caused to it after the issue of this Certificate.
- Any corrections | changes | erasing | modifications | unauthorized Seal, Stamp or Signature on this Calibration Certificate invalidate this Certificate.
- Unless otherwise specified, the measurement data reported is considered 'As found data - without any adjustments.'
- Decision Rule: No conformity statement (Pass | Fail | Accepted | Rejected) is provided, unless otherwise indicated on report.
- NABL Guideline 133 is adopted for the use of NABL Symbol.
- This report pertains only to the sample | equipment submitted and shall not be reproduced in part or full, without explicit written permission from our Laboratory.
- Only Digital calibration Certificate in soft, valid and prepared as per ISO/IEC 17025:20217 requirement is issued.

### Scope of Accreditation:

Please contact our Laboratory to get a copy of our ISO/IEC 17025: 2017 Accredited Certificate and Scope of Calibration. You may also download the same directly from NABL Website.

|                       |           |                                                          |                     |
|-----------------------|-----------|----------------------------------------------------------|---------------------|
| <b>Calibrated By</b>  |           | <b>Reviewed &amp; Approved By   Authorized Signatory</b> |                     |
| <b>Identification</b> | RPL-TM-02 | <b>Identification</b>                                    | Jayasree Manikandan |

|                            |                          |                             |                                 |
|----------------------------|--------------------------|-----------------------------|---------------------------------|
| <b>ULR Number</b>          | ULR - CC417824000000359F | <b>Certificate Number</b>   | RPL-CF-2024-ET-CSRF-1105-C-2513 |
| <b>Date of Calibration</b> | 15.11.2024               | <b>Recommended due date</b> | 14.11.2025                      |

**For any further information, assistance or feedback; please contact:**

Director | Managing Director

Reva Phoenix Labs &amp; Consultants Private Limited

Reva Phoenix Complex, No.14, 4th Street, Rajarajeswari Nagar,

Madipakkam, Chennai 600 091, Tamil Nadu, India

Enquiry : +91 99406 72352 | Accounts: +91 95000 32931 | Reception: +91 98402 96996 | Management: +91 99403 48159

E mail: enquiry@revaphoenix.com | Website: www.revacalibration.com

**Our Accredited Calibration Laboratories:**

- Acceleration, Speed & Vibration, Acoustics
- Mechanical - Density & Viscosity
- Mechanical - Dimension - Basic Measuring Instrument, Gauge etc.
- Mechanical - Dimension - Precision Instruments
- Electro Technical - Time & Frequency, Simulation, Electrical Equipment, AC & DC
- Fluid Flow, Air & Water - Flow Measuring Devices
- Mechanical - Force, Push Pull, UTM, Durometer, Mobile Force Measuring System
- Mechanical - Hardness Testing Machines
- Mechanical - Impact Testing Machine
- Mechanical - Mass & Volume, Weighing Scale & Balance, Weights
- Optical Metrology, Lux Meter & Lamp
- Mechanical - Pressure Indicating Devices
- Thermal - Temperature, Specific Heat & Humidity
- Torque - Generating | Measuring Devices, Torque Wrench, Torque Sensors

**Calibration Services**

Mechanical | Dimension – All Basic & Precision Equipment | CMM Calibration | Pressure – Upto 4000 Bar | Torque – upto 3000 Nm |  
 Mass E1 | Volume – Micro Pipette | Electro-Technical MFC & Oscilloscope | Thermal – Furnace Mapping & Pyrometer upto 1750°C |  
 Force | Optics | Light Source | Lux Meter (1 to 2,00,000 Lux) | Acceleration | Speed | Tachometer | Rubber Hardness Tester | Fluid  
 Flow | Anemometer | Acoustics | Mechanical Calibration | Rockwell | Brinell | Vickers | Micro-Vickers | UTM Calibration | UTM CTM  
 TTM | Impact Testing Machine Calibration | Force Gauge | Push Pull Gauge | Leeb's Hardness | Portable Hardness Tester |  
 Dynamic Hardness Tester | E1 Calibration | Density & Viscosity | Air Flow Meters | pH Meter | Conductivity Meter | Lab Master  
 Equipment Calibration | Gauss | Reference Magnet | 3 Energy & Power | Liquid Flow Meter | Vibration | Portable Vibration  
 Calibrator | Torque Sensor Upto 2000 Nm

|                       |           |                                                          |                     |
|-----------------------|-----------|----------------------------------------------------------|---------------------|
| <b>Calibrated By</b>  |           | <b>Reviewed &amp; Approved By   Authorized Signatory</b> |                     |
| <b>Identification</b> | RPL-TM-02 | <b>Identification</b>                                    | Jayasree Manikandan |


**- End of Calibration Certificate -**

Page No : 05 of 05